



Effective masking for test ear left

	.125	.25	.5	.75	1	1.5	2	3	4	6	8
AC											
BC			75							85	
IP											

Effective masking for test ear right

	.125	.25	.5	.75	1	1.5	2	3	4	6	8
AC											
BC			80								
IP											

SRT AC PTA Right: 41.7 AC PTA Left: 48.3						
Trans	type	dB	Mask	Aided   Binau	List	
IP-R	HL	40	--		Adult Spondees, Form A	
IP-R	MCL	70	--		Rainbow Passage	
IP-L	HL	50	--		Adult Spondees, Form B	
IP-L	MCL	70	--		Rainbow Passage	

WR AC PTA Right: 41.7 AC PTA Left: 48.3						
Trans	WR	dB	Mask	%	Aided   Bin	List
IP-R	WR1	75		90		NU-6 3A - Ordered
IP-L	WR1	75		90		NU-6 4A - Ordered

Audiogram Notes

AUD 8/9/2023 12:29:41 PM

	Right	Left	Bone R	Bone L	FF1	FF2
HL	○	△	×	□	<	>
UCL	U	U	U	U	U	U

Signed By

Printed On: 8/9/2023 1:06:57 PM

AUD, REM, HIT - Affinity Compact  
 Hardware Version: 4.6  
 Serial number: SN1028126  
 Last calibration date: 7/12/2023  
 Next calibration date: 7/11/2024

Name Address 2 Zip code Fax/Mobile  
 Company Nb. City Country Email  
 Address 1 State Telephone